



## Questions & Answers

### 1. What is a LifeVest?

The LifeVest wearable defibrillator is worn by patients at risk for sudden cardiac arrest (SCA), providing protection during their changing condition and while permanent SCA risk has not been established.

### 2. What does the "Respond" message mean?

Before delivering a treatment shock, the LifeVest tests to see if a patient is conscious by providing the patient an opportunity to press and hold the response buttons to prevent a treatment shock. It is important that only the patient press and hold the response buttons.

### 3. What if the patient has Blue™ gel on their skin?

The LifeVest therapy pads release a Blue™ gel prior to a treatment shock to both improve shock conduction and mitigate burning. The gel should remain on the patient as long as the patient is wearing the LifeVest in case additional treatment shocks are required. If you choose to remove the LifeVest from the patient and monitor the patient with external equipment, the gel can be removed with water.

*Questions & Answers continued on back*

### 4. How long does it take for the LifeVest to treat a ventricular arrhythmia?

After the LifeVest detects a treatable arrhythmia, the time to treatment will be between 25 and 60 seconds depending on the type and rate of the arrhythmia and whether the patient presses the response buttons.

### 5. Can emergency personnel get shocked by the LifeVest?

Yes. No one should touch the patient while a shock is delivered. The LifeVest will warn bystanders with both a siren alert and a voice command stating "electrical shock possible, do not touch patient," or "bystanders do not interfere" before a shock is delivered.

### 6. Can emergency personnel use external defibrillation while the patient is wearing a LifeVest?

The monitor should be disconnected from the electrode belt prior to delivering an external defibrillation shock. The garment and belt do not need to be removed.

### 7. What if the patient describes or feels a vibration coming from the garment?

The vibrations, along with the alerts and voice prompts, are part of the LifeVest consciousness test, which requires the patient to press and hold the response buttons to avoid a shock. It is important that only the patient press and hold the response buttons.

### 8. What LifeVest items should the patient bring with them to the hospital?

If possible, the patient should bring the LifeVest, modern, charger, and extra battery to the hospital. This will allow the patient to download any stored event data from the monitor and charge the battery as required.

**24-hour technical support,  
please call: 800.543.3267**

ZOLL • Pittsburgh, PA 15238  
p 800.543.3267 • f 866.567.7615 • [www.zoll.com](http://www.zoll.com)



Blue™ Gel - Only LifeVest has Blue™ gel.

© 2015 ZOLL Medical Corporation. ZOLL, LifeVest, Blue and the color Blue are trademarks and/or registered trademarks of ZOLL Medical Corporation in the United States and/or other countries.

